

FOREST DEVELOPMENT CORPORATION OF MAHARASHTRA  
 (Govt. of Maharashtra Enterprise)  
 CIN: U45200MH1974SGC017206  
 Regd. Office:- FDCM Bhavan, Hingna Road, Nagpur



OFFICE OF THE DIVISIONAL MANAGER, GOREWADA PROJECT  
 Near Hindustan Lever Godown Square, Mahurzhari Road, Fetri, Nagpur-  
 441501

Phone No. 0712-2667484/2667477 Email- [dmgorewada@gmail.com](mailto:dmgorewada@gmail.com)

No. DM/ WRTC/ Automatic Syringe infuser with fluid warmer / 1280 /2019

Date:22 .02.19

## Quotation Letter

M/S \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Subject:** - SUPPLY OF QUATATION for purchase of **Automatic Syringe infuser with fluid warmer**

**Reference- No. DM/ WRTC/ Automatic Syringe infuser with fluid warmer /1109/2019**  
**Date:15 .01.19**

With reference to above subject cited, the quotations were invited as per letter under reference. However, the response was not as per requirement of Government Rules. Therefore, quotations are again invited from bidder for the purchase of **Automatic Syringe infuser with fluid warmer** .

You are requested that kindly quote your **lowest possible rates** for the supply of **Automatic Syringe infuser with fluid warmer** at Wildlife Rescue Centre, Gorewada, Mourzari Road, Fetri, Nagpur-441501 on or before **02/03/2019 by 2.00 pm**.

Further you are also informed that kindly supply the quotation in the name of **The Divisional Manager, Gorewada Project, Gorewada, Nagpur**

Sr. No.	Name of Items	Quantity	Rate /unit	GST	Total								
1	<b>Automatic Syringe infuser with fluid warmer</b> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td><b>A) Automatic Syringe infuser</b></td> </tr> <tr> <td>Should be compact and should have colorful LCD display</td> </tr> <tr> <td>Should have panel lock feature system.</td> </tr> <tr> <td>Should have built-in flow stop feature.</td> </tr> <tr> <td>Should have 90<sup>0</sup> rotation pole clamp convenient for horizontal bar, vertical IV pole &amp; platform use</td> </tr> <tr> <td>Should have flow rate range from 0.1 to 1200 ml/hour and micro mode from 0.1 to 99.9 ml/hour in 0.1 ml increments with continuous mode of operation</td> </tr> <tr> <td>Should have infusion mode like rate mode, drip mode, time mode, body weight mode</td> </tr> <tr> <td>Should have facility for infusion of 0 to 9999 ml of volume</td> </tr> </table>	<b>A) Automatic Syringe infuser</b>	Should be compact and should have colorful LCD display	Should have panel lock feature system.	Should have built-in flow stop feature.	Should have 90 <sup>0</sup> rotation pole clamp convenient for horizontal bar, vertical IV pole & platform use	Should have flow rate range from 0.1 to 1200 ml/hour and micro mode from 0.1 to 99.9 ml/hour in 0.1 ml increments with continuous mode of operation	Should have infusion mode like rate mode, drip mode, time mode, body weight mode	Should have facility for infusion of 0 to 9999 ml of volume	One			
<b>A) Automatic Syringe infuser</b>													
Should be compact and should have colorful LCD display													
Should have panel lock feature system.													
Should have built-in flow stop feature.													
Should have 90 <sup>0</sup> rotation pole clamp convenient for horizontal bar, vertical IV pole & platform use													
Should have flow rate range from 0.1 to 1200 ml/hour and micro mode from 0.1 to 99.9 ml/hour in 0.1 ml increments with continuous mode of operation													
Should have infusion mode like rate mode, drip mode, time mode, body weight mode													
Should have facility for infusion of 0 to 9999 ml of volume													

