

FOREST DEVELOPMENT CORPORATION OF MAHARASHTRA  
 (Govt. of Maharashtra Enterprise)  
 CIN: U45200MH1974SGC017206  
 Regd. Office:- FDCM Bhavan, Hingna Road, Nagpur



OFFICE OF THE DIVISIONAL MANAGER, GOREWADA PROJECT  
 Near Hindustan Lever Godown Square, Mahurzhari Road, Fetri, Nagpur-  
 441501

Phone No. 0712-2667484/2667477 Email- dmgorwada@gmail.com

No. DM/ WRTC/ Orthropaedic drill machine- electrical / 1285 /2019 Date:22 .02.19

## Quotation Letter

M/S \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Subject:** - SUPPLY OF QUATATION for purchase of **Orthropaedic drill machine- electrical**

**Reference- No. DM/ WRTC/ Orthropaedic drill machine- electrical / 1125/2019 Date:17 .01.19**

With reference to above subject cited, the quotations were invited as per letter under reference. However, the response was not as per requirement of Government Rules. Therefore, quotations are again invited from bidder for the purchase of **Orthropaedic drill machine- electrical**.

You are requested that kindly quote your **lowest possible rates** for the supply of **Orthropaedic drill machine- electrical** at Wildlife Rescue Centre, Gorewada, Mourzari Road, Fetri, Nagpur-441501 on or before **02/03/2019 by 2.00 pm**.


Further you are also informed that kindly supply the quotation in the name of **The Divisional Manager, Gorewada Project, Gorewada, Nagpur**.

Sr. No.	Name of Items	Quantity	Rate /unit	GST	Total										
1	<b>Orthropaedic drill machine- electrical</b> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td><b>Orthropaedic drill machine- electrical</b></td> </tr> <tr> <td>Should have motor unit, sterilizable shaft and hand pieces.</td> </tr> <tr> <td>Should have modular system consist of stand, Base plate, motor, ELCB, MCB, Foot switch.</td> </tr> <tr> <td>Should have minimum one Cannulated drilling hand piece with chuck and key.</td> </tr> <tr> <td>Should have minimum one sagital saw</td> </tr> <tr> <td>Should have at least three blades for sagital saw (1.19 mm x 13 mm, 1.19 mm x 19 mm &amp; 1.19 mm x 25 mm)</td> </tr> <tr> <td>Should have minimum one roaming hand piece with low RPM</td> </tr> <tr> <td>Should have minimum one sterilizable flexible shaft</td> </tr> <tr> <td>Should have bided guide wire</td> </tr> <tr> <td>Should have flexible reamer fixed tip 0.8mm</td> </tr> </table>	<b>Orthropaedic drill machine- electrical</b>	Should have motor unit, sterilizable shaft and hand pieces.	Should have modular system consist of stand, Base plate, motor, ELCB, MCB, Foot switch.	Should have minimum one Cannulated drilling hand piece with chuck and key.	Should have minimum one sagital saw	Should have at least three blades for sagital saw (1.19 mm x 13 mm, 1.19 mm x 19 mm & 1.19 mm x 25 mm)	Should have minimum one roaming hand piece with low RPM	Should have minimum one sterilizable flexible shaft	Should have bided guide wire	Should have flexible reamer fixed tip 0.8mm	<b>One</b>			
<b>Orthropaedic drill machine- electrical</b>															
Should have motor unit, sterilizable shaft and hand pieces.															
Should have modular system consist of stand, Base plate, motor, ELCB, MCB, Foot switch.															
Should have minimum one Cannulated drilling hand piece with chuck and key.															
Should have minimum one sagital saw															
Should have at least three blades for sagital saw (1.19 mm x 13 mm, 1.19 mm x 19 mm & 1.19 mm x 25 mm)															
Should have minimum one roaming hand piece with low RPM															
Should have minimum one sterilizable flexible shaft															
Should have bided guide wire															
Should have flexible reamer fixed tip 0.8mm															

Should have flexible reaming stem with removable small & large bits				
Should have reaming bits of diameter 8.5mm, 9.0mm, 9.5mm,10.mm,10.5mm,11.0mm,11.5mm, 12.0mm,12.5mm,13.0mm,				
Should be provided with container for keeping the parts of equipment.				

**Terms & Conditions:**

1. **Those bidder who have submitted quotations earlier will be considered during this advertisement. However, they are free to apply again.**
2. The quotations should be in a sealed envelope mentioning **“Quotations for supply of Orthopaedic drill machine- electrical”** at Wildlife Rescue Centre, Gorewada, Mourzari Road, Fetri, Nagpur-441501 .
3. Mention your GST No. on quotation, otherwise quotation will be rejected.
4. The amount of GST should be mentioned separately.
5. Sealed quotations should reach to the Office of **The Divisional Manager, Gorewada Project, Gorewada, Nagpur** within given stipulated time period on working days and office hours. The quotations received after due date/days shall be rejected.
6. Delivery F.O.R.at **The Divisional Manager, Gorewada Project, Gorewada, Nagpur**
7. This office has right to accept or reject any quotation without prior notice to the concerned.
8. The quantity may increase or decrease.
9. This office has right to reject material at time of delivery if not having the specifications as mentioned in the quotation letter.
10. The brochure/ leaflet of the product must be attached with quotation letter.
11. Supplier has to be given installation and demonstration of instrument / equipment.

  
**Divisional Manager,**  
**Gorewada Project, Gorewada,**  
**Nagpur**