



OFFICE OF THE DIVISIONAL MANAGER, GOREWADA PROJECT
 Near Hindustan Lever Godown Square, Mahurzhari Road, Fetri, Nagpur-
 441501

Phone No. 0712-2667484/2667477 Email- dmgorewada@gmail.com

No. **DM/ WRTC/ Orthropaedic drill machine- electrical / 1125** /2019 Date:17.01.19

Quotation Letter

M/S _____

Subject: - SUPPLY OF QUATATION for purchase of Orthropaedic drill machine- electrical

With reference to above subject cited you are requested that kindly quote your **lowest possible rates** for the supply of **Orthropaedic drill machine- electrical** at Wildlife Rescue Centre, Gorewada, Mourzari Road, Fetri, Nagpur-441501 on or before **30/01/2019 by 5.00 pm.**


Further you are also informed that kindly supply the quotation in the name of **The Divisional Manager, Gorewada Project, Gorewada, Nagpur.**

Sr. No.	Name of Items	Quantity	Rate /unit	GST	Total													
1	<p>Orthropaedic drill machine- electrical</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="padding: 2px;">Orthropaedic drill machine- electrical</td> </tr> <tr> <td style="padding: 2px;">Should have motor unit, sterilizable shaft and hand pieces.</td> </tr> <tr> <td style="padding: 2px;">Should have modular system consist of stand, Base plate, motor, ELCB, MCB, Foot switch.</td> </tr> <tr> <td style="padding: 2px;">Should have minimum one Cannulated drilling hand piece with chuck and key.</td> </tr> <tr> <td style="padding: 2px;">Should have minimum one sagital saw</td> </tr> <tr> <td style="padding: 2px;">Should have at least three blades for sagital saw (1.19 mm x 13 mm, 1.19 mm x 19 mm & 1.19 mm x 25 mm)</td> </tr> <tr> <td style="padding: 2px;">Should have minimum one roaming hand piece with low RPM</td> </tr> <tr> <td style="padding: 2px;">Should have minimum one sterilizable flexible shaft</td> </tr> <tr> <td style="padding: 2px;">Should have bided guide wire</td> </tr> <tr> <td style="padding: 2px;">Should have flexible reamer fixed tip 0.8mm</td> </tr> <tr> <td style="padding: 2px;">Should have flexible reaming stem with removable small & large bits</td> </tr> <tr> <td style="padding: 2px;">Should have reaming bits of diameter 8.5mm, 9.0mm, 9.5mm,10.mm,10.5mm,11.0mm,11.5mm, 12.0mm,12.5mm,13.0mm,</td> </tr> <tr> <td style="padding: 2px;">Should be provided with container for keeping the</td> </tr> </table>	Orthropaedic drill machine- electrical	Should have motor unit, sterilizable shaft and hand pieces.	Should have modular system consist of stand, Base plate, motor, ELCB, MCB, Foot switch.	Should have minimum one Cannulated drilling hand piece with chuck and key.	Should have minimum one sagital saw	Should have at least three blades for sagital saw (1.19 mm x 13 mm, 1.19 mm x 19 mm & 1.19 mm x 25 mm)	Should have minimum one roaming hand piece with low RPM	Should have minimum one sterilizable flexible shaft	Should have bided guide wire	Should have flexible reamer fixed tip 0.8mm	Should have flexible reaming stem with removable small & large bits	Should have reaming bits of diameter 8.5mm, 9.0mm, 9.5mm,10.mm,10.5mm,11.0mm,11.5mm, 12.0mm,12.5mm,13.0mm,	Should be provided with container for keeping the	One			
Orthropaedic drill machine- electrical																		
Should have motor unit, sterilizable shaft and hand pieces.																		
Should have modular system consist of stand, Base plate, motor, ELCB, MCB, Foot switch.																		
Should have minimum one Cannulated drilling hand piece with chuck and key.																		
Should have minimum one sagital saw																		
Should have at least three blades for sagital saw (1.19 mm x 13 mm, 1.19 mm x 19 mm & 1.19 mm x 25 mm)																		
Should have minimum one roaming hand piece with low RPM																		
Should have minimum one sterilizable flexible shaft																		
Should have bided guide wire																		
Should have flexible reamer fixed tip 0.8mm																		
Should have flexible reaming stem with removable small & large bits																		
Should have reaming bits of diameter 8.5mm, 9.0mm, 9.5mm,10.mm,10.5mm,11.0mm,11.5mm, 12.0mm,12.5mm,13.0mm,																		
Should be provided with container for keeping the																		

parts of equipment.				
---------------------	--	--	--	--

Terms & Conditions:

1. The quotations should be in a sealed envelope mentioning “**Quotations for supply of Orthopaedic drill machine- electrical**” at Wildlife Rescue Centre, Gorewada, Mourzari Road, Fetri, Nagpur-441501 .
2. Mention your GST No. on quotation, otherwise quotation will be rejected.
3. The amount of GST should be mentioned separately.
4. Sealed quotations should reach to the Office of **The Divisional Manager, Gorewada Project, Gorewada, Nagpur** within given stipulated time period on working days and office hours. The quotations received after due date/days shall be rejected.
5. Delivery F.O.R.at **The Divisional Manager, Gorewada Project, Gorewada, Nagpur**
6. This office has right to accept or reject any quotation without prior notice to the concerned.
7. The quantity may increase or decrease.
8. This office has right to reject material at time of delivery if not having the specifications as mentioned in the quotation letter.
9. The brochure/ leaflet of the product must be attached with quotation letter.
10. Supplier has to be given installation and demonstration of instrument / equipment.


**Divisional Manager,
Gorewada Project, Gorewada,
Nagpur**