

FOREST DEVELOPMENT CORPORATION OF MAHARASHTRA

(Govt. of Maharashtra Enterprise)

CIN: U45200MH1974SGC017206

Regd. Office:- FDCM Bhavan, Hingna Road, Nagpur



OFFICE OF THE DIVISIONAL MANAGER, GOREWADA PROJECT
Near Hindustan Lever Godown Square, Mahurzhari Road, Fetri, Nagpur-441501
Phone No. 0712-2667484/2667477 Email- dmgorewada@gmail.com

No. DM/ WRTC/ ~~730~~ / 2019

Date: 13.09.2019

Quotation Letter

M/S _____

Subject: - Quotation for Supply and Installation of **Multipara Patient Monitor** at Wildlife Research and Training Centre, Gorewada, Nagpur

With reference to above subject cited you are requested that kindly quote your **lowest possible rates** for the Supply and Installation of **Multipara Patient Monitor** at Wildlife Research and Training Centre, Gorewada, Nagpur on or before **23 /09/2019 by 3.00 pm.**

Further you are also informed that kindly supply the quotation in the name of 'The **Divisional Manager, Gorewada Project, Gorewada, Nagpur**

Sr. No.	Name of Items	Quantity Required	Rate/unit	GST per unit	Total per unit
1	Supply and Installation of Multipara Patient Monitor Technical Specifications: 1) It should have data storage capacity upto atleast 360 hours. 2) It should have facility for printing of data, waveform, trend tables and trend diagrams. 3) It should have multi-language facility. 4) It should have facility for WAN communication and wired or wireless facility and should be HL7 system compatible. 5) It should have minimum 12.1" Flat panel TFT display. 6) It should have flat mirror TFT LCD with touch type key facility. 7) It should have audible and	1			

	<p>visual alarm and adjustable alarm system.</p> <p>8) The minimum size of the monitor should be 12" X 11.3" X 6.3"</p> <p>9) The maximum weight should not more than 8.33 lb.</p> <p>10) It should work in 100-240v.</p> <p>11) It should work on rechargeable lithium battery of 11.1V 4000 mAh.</p> <p>12) It should have standard configuration of SpO₂, PR, NIBP, ECG, RESP and TEMP.</p> <p>13) It should have NIBP measurement facility range from atleast 10 to 270 mmHg with different modes such as manual/auto/continuous modes.</p> <p>14) It should have 3-lead or 5-lead ECG facility with lead selection section I, II, III, avR, avL, avF, V and measurement of heart rate from atleast 15 to 300 bpm.</p> <p>15) It should have SpO₂ measurement facility from 0 to 100%, with a minimum accuracy of 70 to 100%.</p> <p>16) It should have facility for recording of pulse with a minimum accuracy from atleast 20 to 254 bpm.</p> <p>17) It should have facility for recording of Respiration from 0 to 150 rpm.</p> <p>18) It should have facility for recording of temperature from 0 to 50⁰C.</p>				
--	---	--	--	--	--

Terms & Conditions:

1. The quotations should be in a sealed envelope mentioning "**Supply and Installation of Multipara Patient Monitor** at Wildlife Rescue Centre, Gorewada, Mahurzari Road, Nagpur".
2. Mention your GST No. on quotation, otherwise quotation will be rejected.

3. Sealed quotations should reach to the Office of **The Divisional Manager, Gorewada Project, Gorewada, Nagpur** within given stipulated time period on working days and office hours.
4. The quotations received after due date/days shall be rejected. Rate should be mentioned including delivery charges, installation and demonstration.
5. This office has right to accept or reject any quotation without prior notice to the concerned.
6. The quantity may increase or decrease.
7. This office has right to reject material at time of delivery if not having the specifications as mentioned in the quotation letter.
8. The brochure/ leaflet of the product must be attached with quotation letter.
9. The amount of GST should be mentioned separately in the quotation letter.



**Divisional Manager,
Gorewada Project, Gorewada,
Nagpur.**

